# J2 | MEDICAL SUPPLY

### N95 Respirator Masks, Honeywell H910 Plus

## **Technical Data Sheet**

#### **KEY FEATURES**

- 95% particle filtration efficiency
- 99.9% bacterial and viral efficiency
- Silicon and latex free
- Polypropylene Nonwoven Inner and Outer Layers
- Polypropylene/Steel Composite nose clip, without metal exposure for increased safety
- Sizes One Size Fits Most

#### MATERIAL COMPOSITION

- Outer layer Polypropylene Nonwoven: Initially filter and block major particles.
- Filter layer 1 Polypropylene Nonwoven: Absorb Large-sized particles and block fine particles.
- Filter layer 2 Polypropylene Nonwoven (Anti-counterfeit layer): Absorb medium-sized particles and block fine particles.
- Inner layer Polypropylene Nonwoven: Fit the skin and make you feel more comfortable.
- Nose pad Polyurethane Foam: Have a high degree of fitness, mitigate the risk of leakage, reduce the pressure on the nose bridge and thus make you more comfortable.
- Head/ ear band Polyester (latex-free): Feels like cotton. Less likely to cause skin allergy, safer than rubber band.
- Nose clip Polypropylene/steel composite (Built-in type): Bending resistance capacity is 1.5 times of the one of general aluminum clip to ensure more safety. Without metal exposure, more safety.



### **INTENDED USAGE**

- Failure to follow all instructions and limitations concerning the use of the respirator and/or failure to wear the respirator for the whole duration of exposure to contaminants can seriously reduce the performance characteristics of the respirator and lead to illness, injury or death.
- Before Occupational Use of this respirator, a written respiratory protection program must be implemented meeting all the local government requirements. In United Stated, employers must comply with OSHA 29 CFR 1910.134 which includes medical evaluation, training, and fit testing.
- Discard the respirator and replace with a new one if: excessive clogging of the respirator causes breathing difficulty, the respirator becomes damaged, Filtering facepiece respirators are to be inspected prior to each use to assure there are no holes in the breathing zone, especially around weld attachment of headbands and no damage has occurred.
- Visually inspect the filtering facepiece prior to each use to assure that there are no holes in the breathing zone, especially
  around the weld attachment of the headbands to insure no damage has occurred. Discard respirator if any visual signs of
  holes or damage.
- Leave the contaminated area if dizziness, irritation or other distress occurs.

For More Information: 855.615.8633 Orders@J2MedicalSupply.com



### N95 Respirator Masks, Honeywell H910 Plus

Technical Data Sheet - (Continued)

#### TIME USE LIMITATION

- Do NOT use the respirator or enter or stay in a contaminated area under the following circumstances:
- atmosphere contains less than 19.5% oxygen
- atmosphere contains oil aerosols
- · for protection against gases or vapors
- contaminants or their concentrations are unknown or immediately dangerous to life or health
- concentrations or contaminants exceed maximum use concentrations in applicable OSHA standards or applicable government regulations or
- 10 times the PEL (Permissible Exposure Limit), whichever is lower
- for sandblasting, paint-spray operations, asbestos
- Do NOT modify or misuse the respirator.
- Do NOT use the respirator with beards or other facial hair that interferes with direct contact between the face and the edge of the respirator, or any other conditions that may prevent a good face-seal.

### **SHELF LIFE AND STORAGE**

- The temperature: between minus 20 und 38 Celsius (-20 C -- 38 C).
- The humidity: less than 80%.
- Best before: maximal 3 years after produced.

### **WARNINGS**

This respirator helps protect against certain contaminants. It may not
eliminate the risk of contracting disease or infection. Misuse may result in
sickness or death. Materials which may come into contact with the
wearer's skin could cause allergic reactions to susceptible individuals.



### ORDERING INFORMATION

SKU#	DELIVERY	MASKS/BOX	BOXES/CASE	MASKS/CASE
J2N95RM1C	Quoted Upon Request	50	12	600

Listed on CDC NIOSH approved N95 Particulate Filter Face Piece Respirator List - Approval No. 84A-8480

For More Information: 855.615.8633 Orders@J2MedicalSupply.com

